

စိုက်ပျိုးရေး၊ မွေးမြူရေးနှင့်ဆည်မြောင်းဝန်ကြီးဌာန၊ မွေးမြူရေးဆိုင်ရာဆေးတက္ကသိုလ်

Lot 11 Lab Equipment Specifications

No.	Item & Specifications	Quantity
1	<p>DRY BLOCK HEATER</p> <p>Expected specification</p> <p>Number of blocks: 1</p> <p>Length (15 cm) x Width (10 cm) x Depth (7.5 cm)</p> <p>Weight: 1 kg</p> <p>Voltage: 220 – 240 / 115 / 100 V</p> <p>Power input: 165 W</p> <p>Heat output: ≥ 165 W</p> <p>Temperature display and timer: yes</p> <p>Time setting range: 1 – 3000 min</p>	1
2	<p>Universal centrifuge</p> <p>Expected specification</p> <p>Number of tubes: 12 ml x 12 tubes</p> <p>Max. Speed: 14000 rpm</p> <p>Speed range: 200 – 14000 rpm</p> <p>Power input: 135 – 520 W</p> <p>Dimensions: 40 cm x 40 cm x 50 cm (L x W x D)</p> <p>Weight: 20kg – 50 kg</p> <p>Voltage: 120 – 220V</p> <p>Time setting: from: 01 to 60 minutes</p>	1
3	<p>Heating Magnetic Stirrer (up to 1500 rpm)</p> <p>Expected specification</p> <p>Output of 600W</p> <p>Stirred 5-liter capacity</p> <p>Working dimensions 30 cm x30 cm x 20 cm (L x W x D)</p> <p>Machine size: 40 cm x 35 cm x 30 cm (L x W x D)</p>	1