

စိုက်ပျိုးရေး၊ မွေးမြူရေးနှင့်ဆည်မြောင်းဝန်ကြီးဌာန၊ မွေးမြူရေးဆိုင်ရာဆေးတက္ကသိုလ်

Lot 16 Lab Equipment Specifications

No.	Item	Specifications	Quantity
1.	Digital X Ray Machine	<p>MODEL : JADE-40 SD (or) VENU1717 DR (or) Ultisys 3.5 kW Input Power : 110-240V (Free Voltage Input) 50/60Hz Output Rating : Max 4 kW Max DC output Voltage 40-125 KV Max DC Output Current 10 – 400 mA High frequency 32KV X- ray or Microprocessor controlled High Frequency inverter Pixel pitch – 139 μm Number of pixels 3072 × 3072(17"×17" one Unit) Preview time – 5 -10 Sec Laser imager Csl tethered flat panel Detector or 4343W, wireless. Portable, Csl made by DRGEM Limiting Resolution -3.6lp/ mm Storage capacity 56000HU (or) 140kHU</p>	1 No.